

MONITOR

- Minimum size 12"
- Required signals: ECG 3/5 leads, SpO2, EtCo2 mainstream, minimum 2 Invasive pressures, NIBP;
- Visual and audible alarms,
- complete with all the necessary cables and sensors,
- preferably compliant with the Italian medical devices legislation in force in Italy applied to medical devices.
- Supply: 220 V, 50 Hz